



Mouse anti-Melanosome (HMB45)

Cat. No.: BMS010 (16 ml Ready-to-use)

Instructions for use

Intended Use

This antibody is designed for the specific localization of melanocytic differentiation antigen gp100 in formalin-fixed, paraffin-embedded tissue sections.

Anti-Melanosome antibody is intended for in vitro diagnostic use.

Specifications

Specificity:	Human Melanosome
Clone:	HMB45
Isotype:	Mouse IgG1 kappa
Species reactivity:	Human +, others not tested

Summary and explanation

Diagnosis of metastatic amelanotic melanomas can be difficult. They can easily be mistaken for poorly differentiated carcinomas, sarcomas, or large-cell lymphomas. Primary spindle-cell melanomas can be mistaken as spindle-cell carcinomas or different types of mesenchymal neoplasms.

In the literature HMB45 is described as melanocytic differentiating antigen gp100. It is expressed in normal fetal and neonatal melanocytes and some nevus cells, but not in adult melanocytes or intradermal nevi. Hence, the marker is suitable for detection of melanocytic tumours. Other markers for diagnosis of melanomas are: S-100, Tyrosinase, Melan-A/MART1, and MiTF.

Reagent provided

Cell culture supernatant in buffer solution with carrier protein and preservative for stabilisation in the following format:

Ready-to-use: 16 ml (Cat. No. BMS010)

Dilution of primary antibody

None

Storage and handling

The antibody should be stored at 2-8°C without further dilution.

If necessary, dilutions of the antibody should be done with a suitable antibody dilution buffer (e.g. ZUC025 from Zytomed Systems). The diluted antibody should be stored at 2-8°C after use. Stability of this working solution depends on various parameters and has to be confirmed by appropriate controls.

The antibody provided is stable until the expiry date indicated on the label, if stored at 2-8°C. Do not use product after the expiry date. Positive and negative controls should be run simultaneously with all specimens. If unexpected staining is observed which cannot be explained by variations in laboratory procedures and a problem with the antibody is suspected, contact Zytomed Systems' technical support or your local distributor.

Precautions

Use through qualified personnel only. Wear protective clothing to avoid contact of reagents and specimens with eye, skin and mucous membranes. If reagents or specimens come in contact with sensitive area, wash with large amounts of water.

Microbial contamination of the reagent must be avoided, since otherwise non-specific staining may occur. ProClin300 and sodium azide (NaN₃) are used for stabilisation. Reaction of sodium azide with lead or copper in drainage pipes can result in the formation of highly explosive metallic azides. Discard the antibody solution in a large volume of running water to avoid formation of deposits. A material safety data sheet (MSDS) for the pure substances is available upon request.

Staining procedure

Refer to the following table for conditions specifically recommended for this antibody. Also refer to detection system data sheets for guidance on specific staining protocols or other requirements.

Parameters	Zytomed Systems recommendations
*Pre-treatment:	Optional: Heat Induced Epitope Retrieval (for example in citrate buffer pH 6.0)
*Control tissue	Melanoma
*Working dilution	None
*Incubation time	60 minutes

Quality control

The recommended positive control tissue for this antibody is a melanoma. We recommend carrying out a positive and a negative control with every staining run. Please refer to the instructions of the detection system for guidance on general quality control procedures.

Troubleshooting

If you observe unusual staining or other deviations from the expected results please read these instructions carefully, refer to the instructions of the detection system for relevant information or contact your local distributor.

Expected results

The antibody stains positive in the cytoplasm of malignant melanoma cells in formalin-fixed, paraffin-embedded tissue. The interpretation of the results is solely the responsibility of the user. Any experimental result should be confirmed by a medically established diagnostic procedure.

Limitations of the Procedure

Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining, for example variations in fixation and embedding or the inherent nature of the tissue can cause inconsistent results (Nadji and Morales, 1983). Endogenous peroxidase, pseudoperoxidase activity in erythrocytes or biotin may cause non-specific staining depending on the detection system used. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive results with HRP (horse radish peroxidase) detection systems (Omata *et al*, 1980). Inadequate counterstaining and mounting can influence the interpretation of the results.

Zytomed Systems warrants that the product will meet all requirements described from its shipping date until the expiry date is reached, if the product is stored and utilised as recommended. No additional guarantees can be given. Under no circumstances shall Zytomed System be liable for any damages arising out of the use of the reagent provided.

Performance characteristics

Zytomed Systems has conducted studies to evaluate the performance of the antibody utilising a standard detection system. The product has been found to be sensitive and specific to the antigen of interest with minimal or no cross-reactivity.

Bibliography

Bruggen J, et al. Cancer Immunol Immunother 15:200-205, 1983
Gown AM, et al. Am J Pathol 123:195-203, 1986
Wick MR, et al. J Cutan Pathol 15:201-207, 1988
Nadji M and Morales AR. Ann N.Y. Acad Sci 1983; 420:134-139
Abrahamsen HN, et al. Cancer. 2004; 100:1683-91
Baisden BL, et al. Am J Surg Pathol. 2000; 24:1140-6

DeLellis, R., et al. Semin Oncol 14:173-192, 1987
Ordonez NG, et al. Am J Clin Pathol 90:385:390,1988
Yaziji H, Gown AM, Int J Surg Pathol, 11:11-15, 2003
Omata M et al Am J Clin Pathol 1980; 73: 626-632
Vaggelli L, et al. Tumori. 2000; 86:346-8



www.zytomed-systems.de
Zytomed Systems GmbH • Anhaltinerstraße 16 •
14163 Berlin, Germany • Tel: (+49) 30-804 984
990

Explanations of the symbols on the product label

Symbols are used in accordance with ISO 15223-1. Further symbols on the product label might be:



GSH07: Warning / Attention

For Research Use Only